



OFFICE OF THE CHIEF DISTRICT MEDICAL OFFICER, BHADRAK

Ph. No. 06784-251866, 219033 Fax – 06784-251855

E-mail : [edmobdk@rediffmail.com](mailto:edmobdk@rediffmail.com), [dpmubha@ori.n](mailto:dpmubha@ori.n)



No 1649 /BBO Dated, Bhadrak the 20 th April, 2011

Sealed quotation are invited from the authorized suppliers/Local firms having the eligibility criteria's such a valid VAT clearance certificate, TAN No and other required papers as per rules for supply of Consumable articles & Instruments of Red-Cross Blood Bank for the year 2011-12.

The sealed quotation should reach to the undersigned through Regd. Post/Courier Service up to 1 PM within on 05-05-11, which will be opened on the same day at 4 PM.

For more details of the items for quotation & specification please log on to [www.bhadrak.nic.in](http://www.bhadrak.nic.in) or inquiry from the notice board of the undersigned. The authority reserves right to reject all ore any of the quotations without any reason thereof.

Sd/- P.K.Beuria  
Chief District Medical Officer,  
Bhadrak.

Memo No 1700 Dated 20/4/11

Copy to Notice Board of all Wings for in formation and wide publication.

Copy forwarded to the D.I.O.NIC, Bhadrak for information with request for wide publication in your web site.

  
Chief District Medical Officer,  
Bhadrak

2.  
20/4/11

**LIST OF BLOOD BANK COMSUMABLE ITEMS. FOR 2011-2012**

Sl. No	Name of the Consumable articles	Specification	Quantity
1	Anti-A group	10ml Vial	each
2	Anti-B group	10ml Vial	each
3	Anti-Rh D	10ml Vial	each
4	Australia Antigen Test kit (ELISA)	96 test	Each box
5	Australia Antigen Rapid Card/Strip test	Single	Each
6	Syphilis Rapid Strip test kit	Single	Each
7	Anti Human Globulin Serum	05 ml vial	Each
8	Bovine Albumin 22%	10 ml vial	Each
9	Anti H	05ml vial	Each
10	Anti A1	05ml vial	Each
11	Anti A+B	05m. ial	Each
12	HIV Elisa test kit	96 test	Each box
13	HIV Rapid Card test kit	Single	Each
14	HCV Rapid Card test kit	Single	Each
15	HCV ELISA test kit	96 test	Each box
16	Malaria antigen Rapid Card	Single	Each
17	Papa in Enzyme	10 ml vial	Each
18	Blood Collection bag CPD-A1-350ml	Single	Each
19	Blood Collection bag CPD-A1-100ml	Single	Each
20	Glass marking pencil/Pen (Red/Black)	Single	Each
21	Micro pore surgical paper tape 2.5 cm x 9.1 cm	Per Roll	Each
22	Pasture pipette with rubber tit	One dorzen	Each pkt
23	Absorbent cotton wool I.P.500gms	Per Roll 500gm.	Each
24	Micro tips for micro pipette capacity 10 to 200 ul	1000 pices	Each pkt
25	Medicated adhesive strip	100 pices	Each pkt
26	Disposable latex gloves medium size (Lab use)	Per pair	Each
27	Disposable surgical mask	Single	Each
28	Disposable head cap	Single	Each
29	Disposable syringe 2ml with 23G needle (with locking device)	Single	Each.
30	Glass test tube (Borosil) 12 cmx x 100 mm	Single	Each
31	Glass slides (Dimension: 25.4 mmx x 76.2 mm and thickness 1.35 mm)	Per pkt (50)	Each pkt
32	Disposable needle 23G	box (100)	Each box
33	Phenyl white -5 lit packing	Per jar (05lit)	Each jar
34	Skin preparation spay containing Benzalkjonium Chloride with Isopropyl Alcohol 250ml	Per bottle (200ml)	Each
35	Wafers.	10ml vial	Each
36	Dizo water - (5lit Jar)		Each

  
 20/4/11  
 Chief District Medical Officer,  
 Bhadrak.

LIST OF BLOOD BANK INSTRUMENTS. FOR 2011-2012

- 1 Test Tube stand (Alumni) *Sach*  
 2 Digital Analects Balance-CPA-200grams

1	Capacity	200grams
2	Readability	0.01 grams
3	Linearity	Plus minus 0.002 grams
4	Reoridycuvukutt	Plus minus 0.001 grams
5	Dimension	To be declared
6	Stainless steel path	
7	Features	Percentage weighing, counting tar (0-200 grams) auto calibration with built in masses.
8		The balance should be supplied withg graft shield.
9	Power supply	220/240 volts.59 cycles,single phase.
10		The balance should be supplied with graft shield.The equipment should be suitable for 0to 40 C at 95% ambient condition.
11		Complete technical specifications, illustrative technical terature/leaflet etc shall be enclosed along with the offer indicating the model quoted.
12		Complete and satisfactory type test certificate as per T/E specification shall be submitted at

- 3 Auto Clave. *Sach*

Sl. No	Particulars	Minimum Specifications
1	Features	Vertical type. Inner chamber and lid made of stainless steel. Radial locking system. Pressure gauge thermometer safety valve, release valve, water level indicator, stainless steel basket, water outlet, mains indicating lamp, paddle-lifting device, automatic low water cut-off device.
2	(a) Capacity	90 litres, 60lit, 40lit
	(b) Pressure	1 to 1.5 kg per sq.cm
3	Temperature	110° C and 140° C-adjustable in steps of 1° C, digital type thermometer with timer.
4	Power supply	230+/-10 volts.50Hz single phase.
5		Installation, commissioning will be the responsibility of the supplier .
6	SURFACE TREATMENT AND POWER COATING	All the mild steel section shall be degreased, derusted, phosphate powder coated.
7	STANDARD	The apparatus should confirm to IS 14345:1996 with latest amendments or equivalent international standards covering, Markings, Safety requirements with recommendations of safe operations

*[Signature]*  
 Chief District Medical Officer  
~~S. V. S.~~ Bhadrak

*2.*  
 2014 11